



Parts List		
ITEM	QTY	DESCRIPTION
1	1	Cabinet painted steel
2	1	Display
3	1	Mirror box, Black anodized aluminium
4	1	Process gas in, OD 6mm or 1/4"
5	1	Process gas out, OD 8mm welded
6	1	Cell SS AISI 316L
7	1	Myers hubs 3/4", 4-20 mA outputs, relay outputs
8	1	Myers hubs 3/4", power cable, 4-20mA inputs
9	1	Myers hubs 3/4", Ethernet cable (optional)
10	1	Myers hubs 3/4", RS232 cable
11	1	Purge fittings (optional)

Ant.	Pos.	Emne	Prod. std.	Materiale	Kg/stk.	Merknad
A		Updated drawing	11.02.2009 SEF			
Rev.		Change note	Date/sign.	Rev.	Change note	Date/sign.
Date:		Design/drawn:	Approved:	Scale:	■ norsk elektro optikk ■ Norsk elektro optikk A/S Solheimveien 62 A P.b. 384 N-1471 Lørenskog	
06.12.2007		Svein-Erik		Type of project: 	Replacement for:	Replacement by:
LaserGas II MP Cabinet CSA, Class 1 Div 2 version					14411A	Sheet: 1 / 2
Reference:		Calculation:		R:\Illustrasjon og test tegninger\Multipas LGII ATEX DNV\14411.idw		